

2625 Patton Road
Roseville, MN 55113
800-322-0956
www.ablenetinc.com

This declaration of conformity is issued under the sole responsibility of the manufacturer. The device covered by the present declaration is in conformity with all regulations below, including compliance with related Essential Requirements.

Object of the declaration:

Product Name:	Quicktalker 7
Product Model Designator:	Quicktalker 7
Product Part Number(s):	10003501
Basic UDI-DI <required for MDR>	00850011150139
Control Indicator:	2020W20 thru 2025W20
Global Medical Device Nomenclature Code (GMDN) and Description	
Product Options/Accessories:	N/A

2625 Patton Road
Roseville, MN 55113
800-322-0956
www.ablenetinc.com


The object of the declaration described above is in conformity with the following regulations:

EU Regulation	Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices
Device Classification:	Class I based on Annex VIII and Rule 13
Conformity Assessment Path	<i>Not Applicable – Class I device</i>
Name/Address/ID of Notified Body:	<i>Not Applicable – Class I device</i>
Standards and Common Specifications	<p>The products listed above have been tested in a typical configuration as described in the Manufacturer's accompanying documentation and are compliant with the product standards listed below.</p> <p><i>EN 60601-1-2:2001 + A1:2006, Class B for Emissions, Immunity for Non Life-Supporting Equipment Risk Analysis per EN ISO 14971:2007</i></p>

2625 Patton Road
Roseville, MN 55113
800-322-0956
www.ablenetinc.com

Additional information:

EU Authorized Representative:	<i>EUCEREP Roald Dahllaan 33 5629MC – Eindhoven The Netherlands</i>
--------------------------------------	---

Signature (signed for and on behalf of AbleNet): 	Date of Issue: <DD Month YYYY> 13 May 2020
Printed Name: Joe Volp	Place of Issue: AbleNet Inc – Roseville, MN
Title: Director of Marketing	Document Number: DoC_10003501_Quicktalker 7_0513320